



ATEMSCHUTZMASKE YUNYIFU FFP2 NR

YUNYIFU FFP2 NR Atemschutzmasken filtern Aerosole und lassen dadurch so gut wie keine Viren hindurch. Die medizinische Filterpartikelmasken sind aus hochwertigem, festem Stoff gefertigt und an den Rändern verstärkt. Sie sind praktisch, leicht und bequem beim Tragen. Durch ihr Design liegen sie eng an, ohne dabei das Atmen zu erschweren.



Sie sind praktisch, leicht und bequem beim Tragen. Durch ihr Design liegen sie eng an, ohne dabei das Atmen zu erschweren.

Eine **YUNYIFU FFP2 NR Maske** erreicht eine Filterwirkung von > 95 Prozent der in der Luft befindlichen nicht öligen Partikel.

KATEGORIE

- Persönliche Schutzausrüstung
- Kategorie 3
- EN149: 2001 + A1: 2009 FFP2 NR
- 10er Packung / einzeln in Polybag
- in vielen Farben erhältlich

BESTELLUNG

IFMS MED GmbH

Seeböckgasse 59/Tür 2.7
1160 Wien

bestellung@ifms-med.at

www.ifms-med.at

TRAGANLEITUNG



TECHNISCHE DATEN

Filterleistung: >95% (FFP2)

Schichten: 5

**Schaumstoff-Einsatz
am Nasenbügel für einen
erhöhten Tragekomfort!**

EU TYPE EXAMINATION CERTIFICATE

Certificate No: 2163-PPE-943

Respiratory protective devices, filtering half masks to protect against particles manufactured by

Shenzhen Yunyifu Health Technology Co., Ltd.

13A, cedar building B, no.52, tairan sixth road, tian'an community, shatou street, futian district, shenzhen city, CHINA

are tested and evaluated according to

**EN 149:2001 + A1:2009 Respiratory Protective Devices -
Filtering Half Masks to Protect Against Particles -
Requirements, Testing, Marking**

Based on the type examination conducted with the evaluation of test reports, technical file according to Personal Protective Equipment Regulation (EU) 2016/425 Annex 5, it is approved that the product meets the requirements of the regulation.

Product Definition

Model: PM-P2

Filtering half mask

Classification: FFP2 NR

Here by the manufacturer is allowed to use notified body number (2163) and can fix CE mark, as shown below, on the Category III product models given above, with;

- Issuing an appropriate EU Declaration of Conformity according to **Personal Protective Equipment Regulation (EU) 2016/425 Annex 9**.
- Ongoing successful performance in fulfilment of the requirements set out in **Personal Protective Equipment Regulation (EU) 2016/425** and harmonised standards, ensured by assessments based on **Annex 7 (Module C2)** or **Annex 8 (Module D)** of the regulation no later than 1 year from the beginning of serial production

This certificate is initially issued on **05/07/2020** and will be valid for 5 years, if there is no change in the relevant harmonised standard affecting the essential health and safety requirements.



Suat KACMAZ
UNIVERSAL CERTIFICATION
Director



E.U. DECLARATION OF CONFORMITY

Manufacturer:

Name: Shenzhen Yunyifu Health Technology Co., Ltd.

Address : 13A Cedar Building B, No. 52, Tairan Sixth Road, Tian 'An Community, Shatou St, Futian District, Shenzhen City, China.

1. PPE Category III: Filtering Half mask. Model PM-P2.
2. This declaration of conformity is issued under the sole responsibility of the manufacturer, and must accompany an EU type examination in addition to a conformity to type certificate based on internal production control and supervised product intervals.
3. Object of the declaration: Filtering Half Mask, White, Model PM-P2.
4. The object of the declaration described in point 4 is in conformity with the relevant Union harmonisation legislation: PPE Regulation EU 2016/425.
5. References to the relevant harmonised standards used, including the date of the standard, or references to the other technical specifications, including the date of the specification, in relation to which conformity is declared: EU 149:2001+A1:2009.
6. The notified body UNIVERSAL SERTİFİKASYON VE GÖZETİM HİZMETLERİ TİCARET LTD. ŞTİ. Necip Fazıl Bulvarı Keyap Sitesi E2 Blok No:44/84 Yukarı Dudullu Ümraniye / İSTANBUL / TÜRKİYE, NB 2163 performed the EU type-examination (Module B) and issued the EU type-examination certificate 2163-PPE-943.
7. The PPE is subject to the conformity assessment procedure: conformity to type based on internal production control plus supervised product checks at random intervals (Module C2) under surveillance of the notified body 2163.

CE

CEO Signature: _____

Hardy-hou

Approved by (Stamped Seal)

Date of Issue: 2020-07-10

